



U.S. Patent No. 7,494,802 B2

Product Insert

Description

Prokera® Clear is a self-retaining biologic corneal bandage. **Prokera Clear** is comprised of a cryopreserved amniotic membrane graft fastened to an ophthalmic conformer. **Prokera Clear** contains a single layer of cryopreserved amniotic membrane featuring a central aperture of 6 mm, with an inner ring diameter of 17.9 mm and an outer ring diameter of 21.6 mm.

Prokera Clear is manufactured, using the CryoTek® process, from donated human birth tissue according to Current Good Tissue Practice (CGTP) and Current Good Manufacturing Practice (CGMP) regulations established by the US Food & Drug Administration (FDA). **Prokera Clear** is stored in a solution of 1:1, v/v Dulbecco's Modified Eagle Medium (DMEM)/Glycerol containing 20 µg/ml Ciprofloxacin and 1.25 µg/ml Amphotericin B. **Prokera Clear** is a Class II medical device cleared by the FDA. **Prokera Clear** is for single use only in one patient by an ophthalmologist or optometrist.

Indications

- **Prokera Clear** is intended for use in eyes in which ocular surface cells are damaged or underlying stroma is inflamed or scarred. Acting as a self-retaining biologic corneal bandage, **Prokera Clear** effectively treats superficial corneal surface diseases by suppressing inflammation, promoting epithelial healing, and avoiding haze.
- **Prokera Clear** is inserted between the eyeball and the eyelid to maintain space in the orbital cavity and to prevent closure or adhesions. Placement of the conformer also enables application of the cryopreserved amniotic membrane to the ocular surface without the need for sutures.

Contraindications

- **Prokera Clear** should not be used in eyes with glaucoma drainage devices or filtering bleb.

Precautions

- Do not use **Prokera Clear** if the device or packaging is damaged – Contact BioTissue immediately if there is any abnormality observed (e.g. device, labeling, shipping, missing information, etc.). Refer to the Customer Feedback section for reporting.
- **Prokera Clear** exposed to controlled room temperature (20°C to 25°C, 68°F to 77°F) for up to 6 hours may be returned to cold temperature storage in accordance with the Storage section as long as the packaging remains unopened and intact.
- Once the packaging is opened, **Prokera Clear** shall either be transplanted or otherwise discarded.
- Do not sterilize or autoclave the product before use.
- Handle **Prokera Clear** with care during insertion.

Warnings

- Do not use on patients with a history of drug reactions to Ciprofloxacin, Amphotericin B, Glycerol, and/or DMEM.
- As with the use of any human tissue, the possibility of infectious agent transmission cannot be completely eliminated although all screening and microbial testing results for this donor were satisfactory for transplantation.
- It is imperative that the device is stored properly until use. For proper storage instructions, refer to the Storage section.
- If the patient cannot close the eyelid, has an incomplete blink or demonstrates any other signs of exposure, address these clinical issues before or at the same time of placing **Prokera Clear**.
- If the patient cannot tolerate wearing **Prokera Clear**, the device should be removed.

Storage

It is the responsibility of the Tissue Dispensing Service, Tissue Distribution Intermediary, and/or End-User Clinician to maintain tissue intended for transplant in appropriate storage conditions prior to further distribution or transplantation.

Upon receipt, ensure the validated time on the shipper has not expired. Remove the product and store accordingly until use.

Location & Temperature	Use After Receipt
-80°C → 4°C (-112°F → 39.2°F) Example: ultra-low temperature freezer, standard freezer, or standard refrigerator	Within the expiration date printed on product packaging (shelf-life is 2 years from date of manufacture)

Insertion and Removal Instructions

- If frozen, allow **Prokera Clear** to sit at controlled room temperature (20°C to 25°C) in its original unopened packaging for at least 5 minutes.
- Peel open the lid and handle the tray using aseptic techniques.
- Discard the storage media contained within the tray.
- Thoroughly rinse **Prokera Clear** with sterile saline solution while inside the tray to remove the storage media and reduce the potential stinging sensation.
- Remove the center retainer piece from the tray to access the **Prokera Clear**.
- Patient may experience a temporary foreign body sensation upon insertion.
- Advise patient to minimize contact with eye once **Prokera Clear** is placed on the ocular surface.


Insertion



1. Apply topical anesthesia



2. Hold the upper eyelid




3. Ask the patient to look down




4. Insert the **Prokera Clear** into the superior fornix



5. Pull lower eyelid down and slide **Prokera Clear** under the lower eyelid



6. Check centration under the slit lamp



7. Apply a tape-tarsorrhaphy over the lid crease (as needed). Apply appropriate medications (as needed)

See Reverse



Removal



1. Apply topical anesthesia



2. Pull the lower eyelid down



3. Lift lower edge of **Prokera Clear** using a cotton swab or blunt forceps



4. Ask the patient to look down



5. Apply gentle pressure on the upper eyelid



6. Slide the **Prokera Clear** out

- Upon dissolution of membrane or completion of treatment, remove **Prokera Clear**.
- Do not leave in the eye longer than 29 days.

Donor Eligibility and Summary of Records

- This tissue was procured from a donor determined to be eligible based on the results of donor screening and testing. Donor-eligibility determination, based on donor screening and testing for relevant communicable disease agents and diseases, has been performed and documented by BioTissue Holdings Inc.
- A blood specimen, drawn within ± 7 days of donation, has undergone serological testing by a laboratory registered with the FDA to perform donor testing and certified to perform such testing on human specimens in accordance with the Clinical Laboratory Improvement Amendments of 1988 (CLIA) and 42 CFR Part 493, or that has met equivalent requirements as determined by the Centers for Medicare and Medicaid Services (CMS). The donor has been tested for the following infectious diseases and results were non-reactive:
 - HIV-1 & HIV-2 Antibody
 - HIV-1 (RNA-NAT)
 - Hepatitis B Surface Antigen (HBsAg)
 - Hepatitis B Core Antibody (HBcAb)
 - Hepatitis B Virus (HBV, DNA-NAT)
 - Hepatitis C Antibody (HCVAb)
 - Hepatitis C Virus (HCV, RNA-NAT)
 - Syphilis (RPR)
 - HTLV I & II Antibody (HTLV I/II Ab)
 - West Nile Virus (WNV, RNA-NAT)
- This tissue has been deemed eligible for transplantation based on acceptable screening, serological and microbial test results.
- Microbial testing of final product has also been performed to ensure no growth of aerobic, anaerobic or fungal cultures.
- A Certificate of Compliance regarding the manufacturing, storage, shipping and tracking controls for BioTissue products is available upon request.

Recipient Records

It is the responsibility of the Tissue Dispensing Service, Tissue Distribution Intermediary, and/or End-User Clinician to maintain records for the purpose of tracing tissue post-transplantation. The transplanting entity should document the disposition (transplanted or discarded) on the Donor and Recipient Information Card (DRI), attach one of the provided product tracking labels to the DRI and mail to BioTissue. Attach the remaining labels in patient and hospital records.

Customer Feedback

Within the United States: Report any customer feedback, including complaint, error or accident notification promptly to BioTissue at (888) 296-8858.

Outside of the United States: Report feedback to your local tissue provider.

Adverse Events

The FDA requires that information be supplied to the product manufacturer for mandatory reporting of adverse events. Possible significant adverse events include microbial infection and transmission of viral disease. **The doctor is responsible for immediately reporting any adverse event potentially attributable to the use of Prokera Clear to BioTissue.**


For Adverse Events, notify via:

Website: www.biotissue.com/complaints

Phone: (888) 296-8858

Fax: (305) 675-3262

Email: Customerfeedback@biotissue.com



Scan Code



Manufactured by BioTissue Holdings Inc.
8305 NW 27th Street, Suite 101, Doral, FL 33122 USA
biotissue.com | (P) 888-296-8858 | (F) 305-412-4429

FEI #: 3009809074
BT-PI-0026, Rev. 03